



Product Service

EU Quality Management System Certificate

Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices, Annex IX Chapter I

Certificate No. V13 120353 0011 Rev. 00

Manufacturer: Gesellschaft zur Forschung, Entwicklung

und Distribution von Diagnostika im

Blutspendewesen mbH

Altenhöferallee 3

60438 Frankfurt am Main GERMANY

GERIVIAINT

SRN Manufacturer - DE-MF-000025295

The quality management system has been evaluated in accordance with Regulation (EU) 2017/746, Annex IX Chapter I with a positive result.

Details on devices covered by the quality management system are described on the following page(s). The report referenced below summarises the results of the assessment and includes reference to relevant CS, harmonised standards and test reports.

The certified quality management system is subject to periodical surveillance.

If class A devices in sterile conditions are covered by this certificate, the audit was limited to the aspects relating to establishing, securing, and maintaining sterile conditions.

If class B or C excluding self-/near-patient-testing, or class C companion diagnostics devices are covered by this certificate, the quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The periodical surveillance includes further assessment of the technical documentation on the basis of representative samples.

If class D devices, class B or C self-/near-patient testing, or class C companion diagnostics devices are covered by this certificate, an EU Technical Documentation Assessment Certificate in accordance with Annex IX Chapter II is required before placing them on the market.

All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:V13 120353 0011 Rev. 00

Report No.: 713384680

Preceding Certificate No.: V10 120353 0001 Rev. 00

 Valid from:
 2025-11-19

 Valid until:
 2029-06-18

Marta Carnielli

Mat clowed

Issue date: 2025-11-19 Head of Certification IVD





Product Service

EU Quality Management System Certificate

Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices, Annex IX Chapter I

Certificate No. V13 120353 0011 Rev. 00

Classification:

Device Group: IVR 0502 - Infectious diseases: Blood transfusion, transplantation

or cell administration

Intended Purpose: See product certificate

The validity of this certificate depends on conditions and/or is limited to the following:

-none-

Revision History:

Description Rev. Dated Report

2025-11-19 713384680 Supplemented: Device(s)/group of

device(s) added

Administrative merge / transfer to

new Certificate Type